

Centrifuge Performance Report

Customer: University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)	Contact: Andra Lahner
Street: 3901 Rainbow	Phone #: 913-574-3018

Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Cond. Filter Cleaned	Timer Accuracy P/F	Lid Latch Safety Interlock	Brake	Speed Check Low		Speed Check Med		Speed Check High		Temperature©	
														Set	Measured	Set	Measured	Set	Measured	Set	Measured
														+/- 2 C=Pass P= Pass F=Fail							
S-5	X-30R	BECKMAN	64426102	P	P	P	P	P	P	P	P	P	P	100	99.3	2500	2501.3	4700	4699.9	21	22.4
S-6	X-30R	BECKMAN	64424100	P	P	P	P	P	P	P	P	P	P	100	100.4	2500	2499.2	4700	4700.0	3	2.1
S-7	X-30R	BECKMAN	64428135	P	P	P	P	P	P	P	P	P	P	100	99.8	2500	2499.4	4700	4700.5	3	2.3
S-8	X-30R	BECKMAN	64428105	P	P	P	P	P	P	P	P	P	P	100	100.2	2500	2500.2	4700	4699.0	22	23.1
DELP G040																					
S-9	5702R	EPPENDORF	5703BM311350	P	P	P	P	P	P	P	P	P	P	100	100.2	2500	2498.1	4400	4387.4	21	22.1

Test equipment used
 Monarch PLT200 Laser Tach 200, S/N: 1041677, EX Date 03/24/2021 Cert#: 00068851

Technician: Jim Cowart
 Date: 1/19/2021

Notes: Next Due 01/2022

WO # _____
 Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088